

ReserveAP™ Phosphatase-Labeled Affinity Purified Antibody To Human IgM (μ)

Produced in Goat

Catalog No. **Size**
4751-1003 **1.0 mg/mL**



DESCRIPTION

Affinity purified antibody isolated from a pool of serum from goats immunized with purified human IgM was labeled with phosphatase using a modified glutaraldehyde procedure (1).

FORM/STORAGE

Liquid. Store refrigerated at 2 - 8°C. Stable for a minimum of 1 year when stored at 2 - 8°C.

STABILIZER AND PRESERVATIVE

Bovine serum albumin (BSA) is added as a protein stabilizer. A proprietary anti-bacterial agent is added to preserve the product. Non-sterile.

ANTIBODY CONCENTRATION

The concentration of affinity purified antibody is 1 mg as determined by UV absorbance at 280 nm.

E/P RATIO

Molar enzyme/antibody protein ratio is from 2.2:1 to 2.9:1.

SPECIFICITY/CROSS REACTIVITY

Tested by gel diffusion and ELISA techniques as applicable. This product reacts specifically with human IgM. Antibodies to human IgM may cross-react with immunoglobulins of other mammalian species if common binding sites are shared. Product has been adsorbed with purified human IgA and IgG. Reactivity has not been characterized to other human immunoglobulins.

SUGGESTED WORKING DILUTIONS

Optimal working concentration should be determined experimentally. Prepare working dilution in TBS or other buffer such as BSA or Milk Diluent/Blocking Solution (See RELATED PRODUCTS) immediately before use. These buffers are not recommended for long term storage. Suggested starting dilutions are as follows.

Note: In many cases, the antibody may be diluted further than indicated.

ELISA:

1/1000 – 1/5000 1.0 μg/mL - 0.2 μg/mL

Blotting:

1/5000- 1/20000 0.2 μg/mL - 0.05 μg/mL

Histo/Cytochemical Procedures:

1/200 – 1/500 5.0 μg/mL - 2.0 μg/mL

PRODUCT SAFETY AND HANDLING

This product is considered non-hazardous as defined by The Hazard Communication Standard (29 CFR 1910.1200). Avoid contact with skin and eyes. In case of contact or spillage, clean with copious amounts of water. Disposal via sanitary sewer.

REFERENCES

1. Voller, A., et al. (1976). A Manual of Clinical Immunology. American Society for Microbiology, 509-510.

RELATED PRODUCTS

BluePhos Microwell Substrate	Cat. No. 50-88-00
FirePhos Microwell Substrate	Cat. No. 50-87-10
pNPP Microwell Substrate	Cat. No. 50-80-00
BCIP/NBT Western Blot Substrate	Cat. No. 50-81-07
PhosphaGLO Chemiluminescent Substrate	Cat. No. 55-60-03
PhosphaGLO Reserve Substrate	Cat. No. 55-60-01

The product listed herein is for research use only and is not intended for use in human or clinical diagnosis. Nothing disclosed herein is to be construed as a recommendation to use this product in violation of any patents. The information presented above is believed to be accurate. However, said information and product are offered without warranty or guarantee since the ultimate conditions of use and the variability of the materials treated are beyond our control. We cannot be responsible for patent infringements or other violations that may occur with the use of this product. No claims beyond replacement of unacceptable material or refund of purchase price shall be allowed. All claims regarding product performance must be made within 30 days following date of delivery.

Limited Use License

The purchase of this product conveys to the buyer the non-transferable right to use the product in research conducted by the buyer. The buyer cannot sell or otherwise transfer this product or materials made by use of this product to a third party or otherwise use this product or materials made with this product for Commercial Purposes without written approval of KPL, Inc. For additional information, please visit www.kpl.com or the KPL catalog. To obtain a license or approval to use this product for purposes other than those permitted above, contact Director of Sales at (301) 948-7755.